

Duotec A/S  
 Herstedøstervej 19  
 2600 Glostrup  
 Tlf. 4345 9188 - Fax 4343 2330  
 email: info@duotec.dk

Kunde: PN Engineering

Tekniker: Bo Storm Ottosen

Detektorinformation			Områdejustering			Kalibreringsdata			Alarmgrænser			
Fabrikat	Type	Serienummer	Kalibreret for	Span	Enhed	Testgas	Nulpunkt	Span	TWA	STEL	Low	High
RAE	P3K	594-916470	ISO	10	ppm	1	0	10	250	500	500	1000

	Gas	Koncentration	Enhed	Tolerance	Certifikat
1	ISO	10	ppm	5%	2937416
2					
3					
4					

10.05.2021

Dato



Underskrift